















Symbol Glossary Definitions

Symbol	Standard Reference	Standard Title	Symbol Title	Explanatory Text
	ISO 15223-1, Clause 5.1.1	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied	Manufacturer	Indicates the medical device manufacturer.
	EN 980, Clause 5.12	Symbols for use in the labelling of medical devices		
	ISO 15223-1, Clause 5.1.2	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Authorized European Representative	Indicates the Authorized representative in the European Community (EC).
	EN 980, Clause 5.13	Symbols for use in the labelling of medical devices.		
	ISO 15223-1, Clause 5.1.4	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Use-by date: YYYY-MM or YYYY-MM-DD	Indicates the date after which the medical device is not to be used.
	EN 980, Clause 5.3	Symbols for use in the labelling of medical devices.		
	ISO 15223-1, Clause 5.1.5	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Batch code	Indicates the manufacturer's batch code so that the batch or lot can be identified.
	EN 980, Clause 5.4	Symbols for use in the labelling of medical devices.		
	ISO 15223-1, Clause 5.1.6	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Catalog number	Indicates the manufacturer's catalog number, or part number, so that the medical device can be identified.
	EN 980, Clause 5.10	Symbols for use in the labelling of medical devices.		
	ISO 15223-1, Clause 5.2.1	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Sterile	Indicates a medical device that has been subjected to a sterilization process.
	EN 980, Clause 5.7	Symbols for use in the labelling of medical devices.		
	ISO 15223-1, Clause 5.2.3	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Sterilized using ethylene oxide	Indicates a medical device that has been sterilized using ethylene oxide.

Symbol Glossary Definitions

Symbol	Standard Reference	Standard Title	Symbol Title	Explanatory Text
	EN 980, Clause 5.8.2	Symbols for use in the labelling of medical devices.		
	ISO 15223-1, Clause 5.2.4	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Sterilized using gamma irradiation	Indicates a medical device that has been sterilized using irradiation.
	EN 980, Clause 5.8.3	Symbols for use in the labelling of medical devices.		
	ISO 15223-1, Clause 5.2.6	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Do not resterilize	Indicates a medical device that is not to be resterilized.
	EN 980, Clause 5.22	Symbols for use in the labelling of medical devices.		
	ISO 15223-1, Clause 5.2.7	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Non-sterile	Indicates a medical device that has not been subjected to a sterilization process.
	EN 980, Clause 5.23	Symbols for use in the labelling of medical devices.		
	ISO 15223-1, Clause 5.2.8	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Do not use if package is damaged	Indicates a medical device that should not be used if the package has been damaged or opened.
	ISO 15223-1, Clause 5.4.2	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Do no reuse	Indicates a medical device that is intended for one use, or for use on a single patient during a single procedure.
	EN 980, Clause 5.2	Symbols for use in the labelling of medical devices.		
	ISO 15223-1, Clause 5.4.4	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Caution, consult accompanying documents	Indicates the need for the user to consult the instructions for use for important cautionary information, such as warning, contraindications, precautions, etc. that cannot be present on the medical device itself.
	EN 980, Clause 5.11	Symbols for use in the labelling of medical devices.		

Symbol Glossary Definitions

Symbol	Standard Reference	Standard Title	Symbol Title	Explanatory Text
	EN 15986, Clause 4.2	Symbol for use in the labelling of medical devices. Requirements for labelling of medical devices containing phthalates.	DEHP di(2-ethylexyl) phthalate	Indicates presence of di (2-ethylhexyl Phthalate (DEHP).

Symbol Glossary Definitions

SYMBOLS NOT FROM STANDARDS

Symbol	Standard Reference	Standard Title	Symbol Title	Explanatory Text
R_x Only	21 CFR 801.15(c)(1)(i)F	Labeling-Medical devices; prominence of required label statements.	Prescription Only	Requires prescription in the United States.
	21 CFR 801.109	Labeling-Prescription devices.		
CE	765/2008/EC 768/2008/EC MDD 93/42/EEC Articles 4,11,12,17, Annex II)	The requirements for accreditation and market surveillance relating to the marketing of products; Medical Device Directive.	CE Marking	Signifies European technical conformity.
0086	765/2008/EC 768/2008/EC MDD 93/42/EEC Articles 4,11,12,17, Annex II)	The requirements for accreditation and market surveillance relating to the marketing of products; Medical Device Directive.	Notified Body Number	Signifies European Notified Body
QTY	21CFR801.5(b)	Labeling-Medical Devices; adequate directions for use	Quantity	Quantity of devices contained inside the commercial packaging
CAT	N/A	N/A	Catalog number	Indicates the manufacturer's catalog number, or part number, so that the medical device can be identified.